





**EMBUILT
TECHNOLOGIES**

SDPA-3

Solar DC Power Analyser



PENDRIVE
UPTO
16GB

FEATURES

- Solar DC : Volt, Amps, Watt**
- Battery DC : Volt, Amps, Watt**
- Load DC : Volt, Amps, Watt**
- Real Time Clock For Data Logging**
- Big Display For Clear View**
- Efficiency Solar-Battery-Load**
- 8 Testing Modes Of Load using through Solar, Battery, or Both**
- Product Testing + Data Logging**
- Auto Increase Serial Number**
- Pass/Fail Facility As Set Limit**
- Log On Connect In Pen-Drive**
- Software Interfacing Facility**
- Internal Memory upto 32K Test**




COMPREHENSIVE TESTING OF ELECTRICAL APPLIANCES



Home Lighting Industry




SOLAR Industry



Automotive Industry



EMBUILT TECHNOLOGIES



TECHNICAL SPECIFICATION

Auxiliary Power	180 - 250 V, 50 Hz
Solar DC i/p Volt	2.0 - 200 V
Solar DC i/p Amps	10mA to 20Amps
Battery i/p volt	2.0 - 200V
Battery i/p Amps	10mA - 20A
Load Power	200V, 10Amps (Up-to 20A DC range also available)
Accuracy	Class 0.2

*** Range Customization also available**




LOG DATA ON PENDRIVE

Excel Report Generation

Sr No	STATUS	Date	Time	V1	A1	Watt1	V2	A2	Watt2	HR-1	HR-2			
1	PASS	27-09-19	12:49:09	39.39	0.3444	13.57	39.39	0.3458	13.62	39.36	0.3438	13.53	99.04	99.34
2	PASS	27-09-19	12:50:11	39.31	0.3433	13.49	39.29	0.3448	13.55	39.26	0.3427	13.45	99.08	99.31
3	PASS	27-09-19	12:51:13	39.27	0.3433	13.48	39.25	0.3437	13.49	39.22	0.3421	13.42	99.22	99.44
4	PASS	27-09-19	12:52:15	39.24	0.3433	13.47	39.21	0.3448	13.52	39.19	0.3427	13.43	99.41	99.33
5	PASS	27-09-19	12:53:17	39.2	0.3433	13.46	39.2	0.3437	13.47	39.17	0.3404	13.33	99.23	98.96
6	PASS	27-09-19	12:54:19	39.19	0.3433	13.45	39.17	0.3437	13.46	39.15	0.3421	13.35	99.25	99.15
7	PASS	27-09-19	12:55:21	39.16	0.3412	13.36	39.15	0.3416	13.37	39.11	0.3399	13.29	99.46	99.30
8	PASS	27-09-19	12:56:23	39.17	0.3433	13.45	39.15	0.3437	13.46	39.12	0.3416	13.36	99.41	99.29
9	PASS	27-09-19	12:57:25	39.14	0.3412	13.35	39.12	0.3416	13.36	39.09	0.3393	13.26	99.35	99.23
10	PASS	27-09-19	12:58:27	39.14	0.3412	13.35	39.13	0.3427	13.41	39.1	0.3404	13.31	99.19	99.28
11	PASS	27-09-19	12:59:29	39.12	0.3412	13.35	39.11	0.3416	13.36	39.09	0.3393	13.26	99.41	99.26
12	PASS	27-09-19	13:00:32	39.13	0.3412	13.35	39.1	0.3427	13.4	39.08	0.3399	13.28	99.23	99.12
13	PASS	27-09-19	13:01:34	39.11	0.3401	13.3	39.08	0.3405	13.31	39.06	0.3387	13.23	99.25	99.41
14	PASS	27-09-19	13:02:36	39.11	0.3401	13.3	39.09	0.3405	13.31	39.06	0.3387	13.21	99.46	99.22
15	PASS	27-09-19	13:03:37	39.11	0.3401	13.3	39.1	0.3405	13.31	39.06	0.3393	13.25	99.41	99.52
16	PASS	27-09-19	13:04:39	39.11	0.3412	13.34	39.09	0.3427	13.39	39.07	0.3399	13.28	99.41	99.12
17	PASS	27-09-19	13:05:42	39.09	0.3401	13.29	39.08	0.3405	13.31	39.05	0.3387	13.23	99.23	99.38
18	PASS	27-09-19	13:06:44	39.11	0.3412	13.34	39.09	0.3416	13.35	39.06	0.3399	13.27	99.25	99.4
19	PASS	27-09-19	13:07:46	39.09	0.339	13.25	39.07	0.3395	13.26	39.04	0.337	13.16	99.46	99.2



SOFTWARE



- Features:**
- Report Generation
 - Data Logger
 - Real Time View
 - Create .CSV file of data




ORDERING CODE

SDPA -3

Solar DC Power Analyser ← → Three Channels



DIMENSION



Developed & Manufactured By :



EMBUILT TECHNOLOGIES

EMBUILT TECHNOLOGIES

A-404, Vaishali Nagar,
Jaipur-302021, Rajasthan, INDIA
Tel. +91-141-2240935 Mob.: +91-7524853539
E-mail info@embuilt.com
Website : <http://www.embuilt.com>



ACCESSORIES



USB cable



Power cable



Manual



Software CD



Patch Cords



Pendrive